

बिहार पुलिस मुख्यालय, पटना
पुनर्निविदा आमंत्रण सूचना सं०-१६/१२-१३

1. विभाग का नाम- पुलिस महानिदेशक का कार्यालय, बिहार, पटना ।
2. विज्ञापनदाता का पदनाम एवं पता- पुलिस महानिरीक्षक के सहायक (क्यू), बिहार, पटना ।
3. निविदा प्राप्ति की तिथि एवं समय- दिनांक- 14.12.2012 15.00 बजे तक ।
4. निविदा खोलने की तिथि एवं समय-दिनांक- 14.12.2012 16.00 बजे ।
5. निविदा प्राप्ति का स्थान- पुलिस महानिदेशक का कार्यालय, बिहार, पटना के कार्यालय भवन के मुख्य द्वार पर रखी निविदा पेटी।
6. कार्य का व्योरा-

S.No	Items name	Specification	Qty
1	Video Spectral Comparator(basic)	An Instrument to compare, magnify, examine and display chance finger prints, document, finger prints, superimposed finger prints, on a monitor using Video Spectrum technique.	01
2	Digestion Apparatus for GCMS	Specification may be obtained from this office. or on website www. biharpolice. bih.nic.in	01
3.	Aniline point apparatus	Consists of thin film tube, Heater and Control Unit . IP2 ISO 2977, DIN51775,FTM 791-3601 Bath Medium: 400 Ml fo heat transfer fluid (355-000-001) mineral oil should be suitable for this application, 6V Lamp with line card; and drive motor Temperature range-365 to + 107.5° F compatible with latest technology.	01
4.	Water purification system	Produce water for microbiology purpose, physical testing/sterilization / hygienic /temperature control. Produces purified water that meets or exceeds CAP, ASTM, NCCLS,BS from portable tap water. Three step purification in one unit. Removes 95-99% of inorganic ions, 99% of dissolved organics (>200 Dalton),bacteria and particulates, Elix module effects final purification using electrode ionization.	02
5.	Ultra violet lamp (Imported German made)	To produce ultra-violet ray with goggle to protect from UV light.	01
6.	Digital Balance	Bright 7 segment LCD display, Built in overload protection, Automatic Shut Off, Application require 0.0001 gram.	01
7.	Automatic Viscera cutting and distillation apparatus	Competent with latest technology.	01
8	Finger print kit with 220V UV 10M-P & Flourescent powder	A Finger kit to develop Finger prints on any surface including those on Multi coloured surface. The kit should have facility of UV Light & contain Flourescent powders.	01
9	Foot print casting kit	A kit which is used in preparing cast of Sunket foot print / tyre print at scene of crime.	01
Note: Before submitting tender please see tender terms & conditions carefully and then verify your tender from this check list:- <ol style="list-style-type: none"> 1- For one or more items technical bid be kept in one envelope. 2- For one or more items financial bid be kept in separate envelope. 3-Certificate of manufacturer or authorization must be enclosed. 4-Earnest money of 25000/- 5-Income tax returns & copy of PAN No. 6-Turn Over 			

निविदा के नियमों, शर्तों की जानकारी के लिए अधोहस्ताक्षरी के कार्यालय में, पुलिस महानिरीक्षक के सहायक(क्यू), बिहार, पटना के दूरभाष सं०-०६१२-२२१५२९५ पर सम्पर्क किया जा सकता है, कार्यालय के सूचना पट तथा वेबसाइट www.prdbihar.gov.in or www.biharpolice.bih.nic.in पर देखा जा सकता है ।

**पुलिस महानिरीक्षक के सहायक(क्यू),
बिहार, पटना ।**

TERMS & CONDITIONS OF RE-TENDER NO. 16 /2012-13

1. Tender of the firms situated inside /out side the State will be accepted with the condition that if their tenders are approved, the firm shall get it self registered with Commercial Tax Department, Bihar before the purchase order is issued.
2. The tender should be submitted in two parts, (1) Technical bid and (2) Financial bid duly sealed in two separate envelopes super scribed as technical bid and financial bid. The rates shall be quoted in the Financial bid. All charges like BST/CST/Service Tax /VAT etc. shall be clearly mentioned but net rate (in figures and words) including all taxes and duties shall be quoted. Vague offers like indicating taxes "as applicable" will be summarily rejected. There shall be no cutting, over writing or correction on the rates. All relevant papers/certificates/specifications etc. of items should be enclosed in technical bid.
3. The technical and financial bids for each item **should be submitted separately in separate envelopes.** The technical and financial bids for more than one items submitted together in the same envelope shall be summarily rejected. If the financial bid is submitted in the same envelope containing the technical bid then also it shall be summarily rejected.
4. The technical and financial bids should be put in two separate sealed envelopes and the envelopes should be marked "Re-Tender No.16 /2012-13 (Technical Bid) and " Re-Tender No.16 /2012-13 (Financial Bid) along with the name and address of the firm. The sealed envelopes containing the technical and the financial bids should be sent in an another sealed envelope which should be marked as "Re-Tender No. 16 /2012-13 ". This envelope should not bear the name and address of the firm.
5. Up-to-date Income Tax returns (Saral) and a photocopy of PAN of the participating firm should be submitted with the technical bid. I.T. return should have stamp and signature of the concerned department, whether filed manually or electronically.
6. The turn over of the firm for the 2011-12 financial year should be at least double the amount of the supply order which is being issued. A copy of the profit and loss a/c of the firm for the 2011-12 financial year certified by a Chartered Accountant should be submitted along with the tender. If the tenderer is authorised dealer and authorised supplier of manufacturing firm, then the certified details of the turnover of authorizing firm will be accepted.
7. If there is some discount in the price of any item, it should be deducted from price itself and should not be quoted separately.
8. If the product is on DGS & D rate contract then a copy of the contract be attached and pricing be done with all taxes including Bihar Sales Tax.
9. All items will be received at Central Clothing Store, Patna, therefore price should be quoted FOR Central Clothing Store, Phulwarisharif, Patna.
10. **Firms will have to deposit a sum of rupees 25000/- (Twenty five thousand) only as an earnest money in the form of postal saving passbook / NSC/Bank draft duly pledged in favour of the undersigned along with the quotations.**
11. Technical bids will be opened 14.12.2012 at 16:00 PM in the office chamber of the undersigned. Representatives of the tenderer may remain present at the time of opening of the technical bid.
12. The technical bid will be opened first and if it is found as per tender conditions, it will be put before Central Purchase Committee. The firm may have to participate in discussions with this committee and / or be required to demonstrate the quoted product. Any paper / document will not be accepted after opening the tender.
13. Successful firm will have to enter into an agreement after depositing a sum of 5% of the total value of the stores as security money in the form of Postal Saving Pass Book/NSC/Bank Guarantee duly pledged in favour of undersigned.
14. The firm will be required to provide satisfactory after sale service.
15. The firm will be required to supply all the items within 30 days from the date of placing order.
16. Payment for the delivered items will be made after receiving OK report from the Inspection Committee.
17. The firm whose quotation is approved shall be issued A/T. Purchase order shall be issued only after the firm enters into an agreement with the undersigned. Terms & Conditions of the agreement shall be intimated in the A/T.
18. Only such manufactures/authorised dealers and authorised suppliers of manufacturing firm who fulfills the above conditions should quote for supply of items.
19. Updated Technology at the time of supply change in condition.
20. Hands on Training of the user group by the firm shall be ensured.
21. D.G. of Police Bihar, Patna reserves the right to reject any or all the quotations partially or fully without assigning any reason thereof.
22. The Quantity indicated may increase or decrease at the time of issuing purchase order.

A I G of Police (Q)
Bihar, Patna.

Specification of Microwave Digestion System For GC-MS

- 1. The system must have a power output of 1600W with a continuous power output. Pulsed units will not be considered.**
- 2. Microwave frequency must be 2450 Mega Hz.**
- 3. Multi-leveled power settings must be available.**
- 4. The microwave cavity should have a Fluoropolymer coating.**
- 5. A corrosion resistant exhaust system must be in place.**
- 6. The door must be impact resistant and have safety interlocks.**
- 7. Microwave unit should have the option of operating via an on board computer or desktop computer.**
- 8. Real time graphics or alphanumeric displays should be available to track the progress of the method in use.**
- 9. The system should have a built-in library of EPA methods available.**
- 10. The software must provide method storage, data collection, real time displays of temperature and pressure, and report writing capabilities.**
- 11. Temperature up to 250⁰c which control should be achieved through a dual system with the use of a single control vessel and a remote IR sensor that monitors individual vessels.**
- 12. Pressure up to 0.1-5mpa (Precision + 0.01 mpa) which is controled by the use of a single control vessel is acceptable.**
- 13. The digestion assembly must have a minimum of 11 prep vessels and a single control vessel.**
- 14. Digestion vessels should have all TFM wetted surfaces.**
- 15. The unit should include a torque station.**
- 16. Installation should include a minimum of a 2 days training on the operation of the unit, routine maintenance and basic application training.**